



# CERTIFICATE OF ACCREDITATION

## The ANSI National Accreditation Board

Hereby attests that

### PROMOCION MEDICA S.A.

Costa del Este, Parque Industrial, Calle 2da Edificio Promed  
Panama

Fulfills the requirements of

### ISO/IEC 17025:2017

In the field of

### CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

A handwritten signature in black ink, appearing to read 'R. Douglas Leonard Jr.', is positioned above a horizontal line.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 25 February 2024  
Certificate Number: AC-2854



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017**

**PROMOCION MEDICA S.A.**

Costa del Este, Parque Industrial, Calle 2da Edificio Promed  
Panama

Epifanía Riley de Rotar

Tel: (+507) 303-3115 E-Mail: [ederotar@promed-sa.com](mailto:ederotar@promed-sa.com)

**CALIBRATION**

Valid to: **February 25, 2024**

Certificate Number: **AC-2854**

**Acoustics and Vibration**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Audiometry Equipment (Audiometers, Optoacoustics Emissions equipment, Impedance meters, sound cameras)	(75 to 106) dB (10 Hz to 20 kHz)	0.47 dB	ANSI/ASA S3.6

**Chemical Quantities**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Potential of Hydrogen-pH <sup>1,3</sup>	4 pH 7 pH 10 pH	0.014 pH 0.015 pH 0.015 pH	Certified Reference Materials - Comparison Method
Conductivity Meters <sup>1,3</sup>	5 µS/cm 100 µS/cm 1 413 µS/cm 100 mS/cm	0.62 µS/cm 2.1 µS/cm 4.9 µS/cm 0.37 mS/cm	Conductivity Certified Materials - Comparison Method

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Energy/Defibrillator <sup>1</sup>	(15 to 360) J	0.017 % of reading + 0.63 J	Fluke Impulse 6000DP Defibrillator Analyzer



ANSI National Accreditation Board

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of pH Meters <sup>1</sup>	(-2 000 to 2 000) mV	0.1 mV	THERMO ELECTRIC ISOCAL 9000

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure/ Blood Pressure Cuff <sup>1</sup>	(60 to 300) mmHg	0.002 % of reading + 0.36 mmHg	DRUCK DPIN610 / GE 2200-A145, Pressure Module 2200-A145
Analytical Balance <sup>1</sup> Resolution: ≥ 0.001 mg ≥ 0.001 mg ≥ 0.001 mg ≥ 0.01 mg	(0 to 100) mg (0 to 22) g (0 to 320) g (0 to 520) g	0.002 mg 0.02 mg 0.083 mg 0.051 mg	Class Weights Mass – OIML Class E2 and F1 for Balance Resolution ≥ 0,1 mg; Comparison Method
Balances / Weighing Instruments <sup>1</sup> Resolution: ≥ 1 mg ≥ 5 mg ≥ 0.01 g ≥ 0.01 g ≥ 0.1 g	(0 to 610) g (0 to 64 100) g (0 to 4 200) g (0 to 10 200) g (0 to 32 200) g	0.021 g 0.011 g 0.019 g 0.13 g 0.26 g	Class Weights Mass – OIML Class E2, F1, M1; Comparison Method
Balances / Scales Floor Scale, Weighing Instruments <sup>1</sup> Resolution: ≥ 0.01 kg ≥ 0.01 kg ≥ 0.02 kg ≥ 0.05 kg ≥ 0.05 kg ≥ 0.2 kg ≥ 0.5 kg	(0 to 150) kg (0 to 300) kg (0 to 600) kg (0 to 500) kg (0 to 1 000) kg (0 to 2 000) kg (0 to 3 000) kg	0.02 kg 0.035 kg 0.06 kg 0.045 kg 0.12 kg 0.47 kg 0.5 kg	Class Weights Mass – OIML Class F1, F2, M1, M2; Comparison Method



ANSI National Accreditation Board

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass: OIML Classes E2, F1, F2, M1, M2 & M3	1 mg	0.002 mg	Weight Set Class E1, E2 Mass Comparators: Mettler Toledo Model XPE56C Mettler Toledo Model XPE505C Mettler Toledo Model XPR2004SC
	2 mg	0.002 mg	
	5 mg	0.002 mg	
	10 mg	0.002 6 mg	
	20 mg	0.003 3 mg	
	50 mg	0.004 mg	
	100 mg	0.005 3 mg	
	200 mg	0.006 7 mg	
	500 mg	0.008 3 mg	
	1 g	0.01 mg	
	2 g	0.013 mg	
	5 g	0.016 mg	
	10 g	0.02 mg	
	20 g	0.026 mg	
	50 g	0.033 mg	
100 g	0.053 mg		
200 g	0.1 mg		
500 g	0.26 mg		
Mass: OIML Classes E2, F1, F2, M1, M2 & M3	1 kg	0.53 mg	Weight Set Class E1, E2 Mass Comparators: Mettler Toledo XPR10003SC Mettler Toledo XPR64003 LD5C
	2 kg	1 mg	
	5 kg	2.7 mg	
	10 kg	5.3 mg	
	20 kg	10 mg	
	25 kg	20 mg	
	50 kg	27 mg	
Piston Volume Devices <sup>1</sup>	1 µL	5 % of indicated volume	Gravimetric Calibration Referenced to Mass Balances: Mettler Toledo XP26PC -Mettler Toledo SAG105 -Mettler Toledo MCP105 (Movil Balance) ISO 8655 Family
	1.25 µL	2 % of indicated volume	
	2 µL	2 % of indicated volume	
	2.5 µL	2 % of indicated volume	
	5 µL	0.8 % of indicated volume	
	10 µL	0.6 % of indicated volume	
	20 µL	0.3 % of indicated volume	
	25 µL	0.3 % of indicated volume	
	30 µL	0.3 % of indicated volume	
	50 µL	0.2 % of indicated volume	
	100 µL	0.3 % of indicated volume	
	150 µL	0.3 % of indicated volume	
	200 µL	0.3 % of indicated volume	
300 µL	0.3 % of indicated volume		



ANSI National Accreditation Board

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Piston Volume Devices <sup>1</sup>	500 µL 600 µL 1 mL 1.2 mL 2 mL 2.5 mL 5 mL 10 mL 12.5 mL 25 mL	0.2 % of indicated volume 0.2 % of indicated volume 0.1 % of indicated volume 0.2 % of indicated volume 0.1 % of indicated volume 0.3 % of indicated volume 0.1 % of indicated volume 0.1 % of indicated volume 0.1 % of indicated volume 0.1 % of indicated volume	Gravimetric Calibration Referenced to Mass Balances: Mettler Toledo XP26PC -Mettler Toledo SAG105 -Mettler Toledo MCP105 (Movil Balance) ISO 8655 Family
Motor Driven Piston Burettes	1 mL 2 mL 5 mL 10 mL 20 mL 25 mL 50 mL 100 mL	0.2 % of indicated volume 0.2 % of indicated volume 0.1 % of indicated volume 0.07 % of indicated volume 0.07 % of indicated volume 0.07 % of indicated volume 0.05 % of indicated volume 0.03 % of indicated volume	Gravimetric Calibration Referenced to Mass Balances, ISO 8655 Family  ISO 8655-3 Piston Burettes
Manual Piston Burettes	1 mL 2 mL 5 mL 10 mL 20 mL 25 mL 50 mL 100 mL	0.2 % of indicated volume 0.2 % of indicated volume 0.1 % of indicated volume 0.1 % of indicated volume 0.07 % of indicated volume 0.07 % of indicated volume 0.07 % of indicated volume 0.07 % of indicated volume	Gravimetric Calibration Referenced to Mass Balances, ISO 8655 Family  ISO 8655-3 Piston Burettes
Piston Dispensers	0,01 mL 0.02 mL 0.05 mL 0.1 mL 0.2 mL 0.5 mL 1 mL 2 mL 5 mL 10 mL 25 mL 50 mL 100 mL 200 mL	0.7 % of indicated volume 0.7 % of indicated volume 0.5 % of indicated volume 0.5 % of indicated volume 0.3 % of indicated volume 0.3 % of indicated volume 0.2 % of indicated volume	Gravimetric Calibration Referenced to Mass Balances, ISO 8655 Family  ISO 8655-5 Dispensers



ANSI National Accreditation Board

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Piston Dilutors	0.05 mL 0.1 mL 0.2 mL 0.5 mL 1 mL 2 mL 5 mL 10 mL 25 mL 50 mL 100 mL	0.6 % of indicated volume 0.5 % of indicated volume 0.3 % of indicated volume 0.3 % of indicated volume 0.2 % of indicated volume	Gravimetric Calibration Referenced to Mass Balances, ISO 8655 Family  ISO8655-4 Dilutors
Laboratory Glassware/ Burettes	1 mL 2 mL 5 mL 6 mL 10 mL 25 mL 30 mL 50 mL 100 mL	0.2 % of indicated volume 0.2 % of indicated volume 0.06 % of indicated volume 0.1 % of indicated volume 0.05 % of indicated volume 0.06 % of indicated volume 0.1 % of indicated volume 0.03 % of indicated volume 0.12 % of indicated volume	Gravimetric Calibration Balances: Mettler Toledo XP26PC Mettler Toledo SAG105 Mettler Toledo XS203S ISO 385 STANDARD
Laboratory Glassware/ Graduated Pipettes	0.1 mL 0.2 mL 0.5 mL 1 mL 2 mL 5 mL 10 mL 20 mL 25 mL 50 mL 100 mL	2 % of indicated volume 2 % of indicated volume 0.7 % of indicated volume 0.2 % of indicated volume 0.14 % of indicated volume 0.07 % of indicated volume 0.05 % of indicated volume 0.035 % of indicated volume 0.03 % of indicated volume 0.02 % of indicated volume 0.015 % of indicated volume	Gravimetric Calibration Balances: Mettler Toledo XP26PC Mettler Toledo SAG105 Mettler Toledo XS203S ISO 835 STANDARD
Laboratory Glassware/ Single Volume Pipettes	0.5 mL 1 mL 2 mL 5 mL 10 mL 20 mL 25 mL 50 mL 100 mL	0.7 % of indicated volume 0.2 % of indicated volume 0.14 % of indicated volume 0.07 % of indicated volume 0.05 % of indicated volume 0.035 % of indicated volume 0.03 % of indicated volume 0.02 % of indicated volume 0.015 % of indicated volume	Gravimetric Calibration Balances: Mettler Toledo XP26PC Mettler Toledo SAG105 Mettler Toledo XS203S ISO 648 STANDARD



ANSI National Accreditation Board

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Laboratory Glassware/ One Mark Volumetric Flasks/ Graduated Flasks	1 mL	1.13 % of indicated volume	Gravimetric Calibration Balances: Mettler Toledo XP26PC Mettler Toledo SAG105 Mettler Toledo XS203S Mettler Toledo Model XPE505C Mettler Toledo Model XPR2004SC Mettler Toledo XPR10003SC Mettler Toledo XPR64003 LD5C ISO 1042 STANDARD
	2 mL	1.13 % of indicated volume	
	5 mL	0.25 % of indicated volume	
	10 mL	0.15 % of indicated volume	
	20 mL	0.13 % of indicated volume	
	25 mL	0.1 % of indicated volume	
	50 mL	0.075 % of indicated volume	
	100 mL	0.05 % of indicated volume	
	200 mL	0.04 % of indicated volume	
	250 mL	0.035 % of indicated volume	
	500 mL	0.03 % of indicated volume	
	700 mL	0.03 % of indicated volume	
	800 mL	0.03 % of indicated volume	
1 000 mL	0.025 % of indicated volume		
2 000 mL	0.025 % of indicated volume		
Test Tubes	5 mL	0.4 % of indicated volume	Gravimetric Calibration ISO 4788 STANDARD
	10 mL	0.2 % of indicated volume	
	25 mL	0.56 % of indicated volume	
	50 mL	0.2 % of indicated volume	
	100 mL	0.2 % of indicated volume	
	200 mL	0.8 % of indicated volume	
	250 mL	0.2 % of indicated volume	
	500 mL	0.4 % of indicated volume	
1 000 mL	0.2 % of indicated volume		
Chemical Glasses	500 mL	0.1 % of indicated volume	Gravimetric Calibration ISO 1042 STANDARD
	600 mL	0.08 % of indicated volume	
	700 mL	0.07 % of indicated volume	
	800 mL	0.06 % of indicated volume	
	1 000 mL	0.07 % of indicated volume	
Pycnometer	25 mL	0.005 % of indicated volume	Gravimetric Calibration ISO 3507 STANDARD
	50 mL	0.005 % of indicated volume	
	100 mL	0.005 % of indicated volume	
Metallic Volumetrics	19 L	0.015 % of indicated volume	Gravimetric Calibration OIML R120
	20 L	0.015 % of indicated volume	

**Photometry and Radiometry**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Wavelength / Spectrophotometer <sup>1</sup>	(279 to 637) nm	0.05 nm	Holmium Oxide Reference Material
Absorbance /Photometric Scale <sup>1</sup>	1 % Au 3 % Au 10 % Au 20 % Au 30 % Au 50 % Au 90 % Au	0.006 Au 0.006 Au 0.003 Au 0.003 Au 0.003 Au 0.002 5 Au 0.002 5 Au	Neutral Density Filters with Different Transmittance Percentages

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature/ Digital Thermometers Direct Indication Thermometers Temperature Data Loggers Bimetallic Thermometers <sup>1</sup>	(-30 to 250) °C	0.05 °C	Digital Thermometer ISOTECH 935-14-95H ISOTECH T100-250-316-9 TESTO 614.024 Bath: INSCO 777; ISOTECH Orion 796 H; ISOTECH Fast Cal
Liquid-in-Glass Thermometers	(-30 to 250) °C	0.1 °C	
Infrared (IR) Thermometers	(30 to 45) °C	0.67 °C	Infrared Blackbody Temperature Calibrator $\epsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$
Temperature Measure/ Incubators, Coolers/Ovens/Circulating Baths/Environmental Chambers	(-80 to 35) °C (35 to 1 000) °C	0.13 °C	Testo Type K probe LMB100 Testo Type K probe LMB

**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Centrifugal Speed <sup>1,2</sup>	(10 to 25) rpm (25 to 100) rpm (100 to 1 000) rpm (1 000 to 93 750) rpm	0.001 % of reading + 0.1 rpm 0.001 % of reading + 0.1 rpm 0.000 5 % of reading + 0.1 rpm 0.000 5 % of reading + 1 rpm	Digital Tachometer Exttech 461995, Digital Tachometers Testo 465 and 470

**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Cardiac Rate/ECG Multi-parameter Monitor <sup>1,2</sup> (Electrical Simulation)	(60 to 300) BPM	1.5 BPM	Patient simulator MPS450
Audiometry Equipment (Audiometers, Optoacoustics Emissions equipment, Impedance meters, sound cameras)	251.2 Hz	0.62 Hz	ANSI/ASA S3.6

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope
2. d = resolution of device under test; rpm = revolutions per minute; BPM = beats per minute.
3. The nominal values listed are approximate.
4. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2854.



R. Douglas Leonard Jr., VP, PILR SBU